

[Klik hier voor de LINK naar de Nederlandse versie](#)

Revision changes

Concentration citrate is added.

25-03-2018: Biobank tracking system is changed to Biobank management system

1. Goal

To describe how citrate plasma (9 ml) must be prepared and stored within the framework of the Parelsnoer Institute for the purposes of the Radboud Biobank, so that:

1. Potential users know how the biomaterial has been handled.
2. The laboratory preparing the biomaterial for storage knows the minimal requirements for the handling of it.

1.1 Scope of application

This procedure is applicable to all employees concerned at the Radboud Biobank and the Radboud Laboratory for Diagnostics.

2. Protocol citrate plasma 9 ml

Collection

Fasting patient	Dependant on the requirements of each Pearl/sub-biobank.
Delivery	In collection tube.
Type of tube	Standard 9 ml tube, 3,2% sodium citrate, no glass tubes unless silicone lined, without gel, without protease inhibitors. Greiner tube 455322 or equivalent.
Temperature	Keep tubes at room temperature until prepared for storage.

Preparation

Time until freezing	Prepare and freeze the collected blood as soon as possible.. Aim: within 2 hours. Maximum: within 4 hours.
Centrifuge	g-force: 2000g Time: 10 minutes. Temperature: At room temperature.

Storage

Register	Register the microtubes in the Biobank management system and attach a cryovial label.
Aliquotation	Store the biomaterial in 0.5 ml aliquots to a maximum of 6 aliquots. If there is biomaterial left over add it to the 6 aliquots.
Type of tube	2 ml polypropylene microtube and screw-top with O-ring. GREI722301UMC and GREI368380UMC or equivalent.
Temperature	-80°C

Miscellaneous

Deviations

Note all deviations from this procedure in the **Biobank management system**.

3. Quality assurance

Review takes place via an internal audit (see the audit plan).

3.1 Performance indicator

95% of the relevant samples have been prepared and stored within the parameters of this procedure.

4. Comments

-

5. Relevant documents

Qdocs, literature, legislation and codes of conduct, website etc.

link	Titel
Qdoc 045014	Radboud Biobank voorschrift RLD (NB alleen bestemd voor medewerkers RLD)
Qdoc 078088	Monstermanagement RLD t.b.v. Radboud Biobank

Wijzigingen bij revisie

Concentratie citraat is toegevoegd.

1. Doel

Vastleggen hoe citraat plasma (9 ml) binnen de kaders van het Parelsnoer Instituut moet worden bewerkt en opgeslagen wanneer dit binnen de Radboud Biobank wordt verzameld, zodat:

1. Potentiële uitnemers weten hoe het materiaal verwerkt is.
2. Het uitvoerend laboratorium weet wat de minimale eisen zijn voor de verwerking van het materiaal.

1.1 Toepassingsgebied

Deze procedure is van toepassing voor alle betrokken medewerkers van de Radboud Biobank en het Radboud Laboratorium voor Diagnostiek.

2. Protocol citraat plasma 9 ml

Afname

Patiënt nuchter	Afhankelijk van afspraken binnen een parel/deelbiobank.
Aanlevering	In afnamebuis.
Type afnamebuis	Standaard 9 ml buis, 3,2% natriumcitraat, geen glazen buis tenzij gesiliconeerd, geen gel, geen proteaseremmers. Greiner buis 455322 of gelijkwaardig.
Temperatuur	Buizen tot bewerking bewaren op kamertemperatuur.

Bewerking

Tijd tot invriezen	Afgenomen bloed zo snel mogelijk bewerken en invriezen. Streeftijd: binnen 2 uur. Maximum: binnen 4 uur.
Centrifugeren	g-kracht: 2000g Tijd: 10 minuten. Temperatuur: Bij kamertemperatuur.

Opslag

Aanmelden	Microbuizen aanmelden in biobankbeheersysteem en voorzien van een cryovioletiket.
Aliquotering	Materiaal opslaan in aliquots van 0,5 ml. Maximaal 6 aliquots van 0,5 ml opslaan. Als er meer materiaal is, dit opslaan door het restant aan de aliquots toe te voegen.
Type buis	2 ml polypropyleen microbuis en schroefdop met O-ring. GREI722301UMC en GREI368380UMC of gelijkwaardig.
Temperatuur	-80°C

Title: SOP citrate plasma 9 ml

Overige

Afwijkingen Elke afwijking van het bovenstaande noteren in het biobankbeheersysteem

3. Kwaliteitsborging

Toetsing vindt plaats via interne audit (zie verder planning Interne audits).

3.1 Prestatie-indicator

95% van de monsters zijn bewerkt volgens de eisen van deze SOP.

4. Opmerkingen

-

5. Relevante documentatie

Qdocs, literatuur, wet en regelgeving, website e.d

link	Titel
Qdoc 045014	Radboud Biobank voorschrift RLD (NB alleen bestemd voor medewerkers RLD)
Qdoc 078088	Monstermanagement RLD t.b.v. Radboud Biobank